

## Foshan Senda Medical Instrument Co.,Ltd

Add: Adress: 5/F, Building F4, Guangdong New Light Source Industrial Base, Luocun town, Nanhai Dist, Foshan City, Guangdong Prov, China

Phone:86 13924878366 Web: http://www.sendadental.com

## **Product Specifications**

Brand: Jaden

Name: High speed air turbine handpiece

Model: SD-603 M4 C

No.	ltem	Technical parameters & performance indicators	Remark
1	Appearance description	Led light, copper body	
2	Head size	Size Φ12.5x14.3	
3	Bearing	Ceramic bearings	
4	Internal structure	The wind wheel assembly is an open type	
5	Water	3-holes water spray	
6	Health system	Anti suction back system	
7	Chuck type	Push button	The chucking force of the fixture on the needles during operation is ≥ 28N
8	Tail design	Four-hole fixed connector	
9	Air pressure	0.22-0.25MPa	
10	Air consumption	30-36L/min	
11	Cartridge balance	≤150µg	
12	Radial runout	≤0.01 mm	
13	Rotational speed	310,000-400,000rpm/min	
14	Cutting force	Radial≥1.1Kg Axia≥2.0Kg	
15	Noise	≤60dB	
16	Other	When the water pressure is 200kpa, the cooling water flow rate is 90-110mL/min	
17	135°C Autoclavable		
18	Warranty one year		

